



EU Quality Management System Certificate (IVDR)

Pursuant to Regulation (EU) 2017/746 on in Vitro Diagnostic Medical Devices,
Annex IX Chapters I and III (Class C and B Devices excluding self-/near-patient-testing and
Companion Diagnostics)

No. V12 058438 0015 Rev. 00

Manufacturer: **AB ANALITICA Srl**
Via Svizzera, 16
35127 Padova
ITALY

SRN Manufacturer: Not available at issuance date of this certificate

The Certification Body of TÜV SÜD Product Service GmbH certifies that the manufacturer has established, documented and implemented a quality management system as described in Article 10 (8) of the Regulation (EU) 2017/746 on in Vitro Diagnostic Medical Devices. Details on devices covered by the quality management system are described on the following page(s).

The Report referenced below summarizes the result of the assessment and includes reference to relevant CS, harmonized standards, audit and test reports. The conformity assessment has been carried out according to Annex IX Chapter I and III of this regulation with a positive result.

The quality management system assessment was accompanied by the assessment of technical documentation for devices selected on a representative basis.

The certified quality management system is subject to periodical surveillance by TÜV SÜD Product Service GmbH. The surveillance assessment includes an assessment of the technical documentation for the device or devices concerned on the basis of further representative samples.

For details and certificate validity see: www.tuvsud.com/ps-cert?q=cert:V12_058438_0015_Rev.00

Report No.: ITA1691857
Valid from: 2022-09-28
Valid until: 2027-09-27

Christoph Dicks
Head of Certification/Notified Body

Issue date: 2022-09-28



EU Quality Management System Certificate (IVDR)

Pursuant to Regulation (EU) 2017/746 on in Vitro Diagnostic Medical Devices, Annex IX Chapters I and III (Class C and B Devices excluding self-/near-patient-testing and Companion Diagnostics)

No. V12 058438 0015 Rev. 00

Classification: C
Device Group: W0105 - INFECTIOUS DISEASES
IVP Code: IVP 3011 - In vitro diagnostic devices which require knowledge regarding molecular biological testing including nucleic acid assays and next generation sequencing (NGS)
Intended Purpose: IVR 0503 - Devices intended to be used to detect the presence of, or exposure to an infectious agent including sexually transmitted agents

Classification: B
Device Group: W0105 - INFECTIOUS DISEASES
Intended Purpose: IVR 0503 - Devices intended to be used to detect the presence of, or exposure to an infectious agent including sexually transmitted agents

Classification: C
Device Group: W0106 - GENETIC TESTING
IVP Code: IVP 3011 - In vitro diagnostic devices which require knowledge regarding molecular biological testing including nucleic acid assays and next generation sequencing (NGS)
Intended Purpose: IVR 0401 - Devices intended to be used in screening/confirmation of congenital/inherited disorders

The validity of this certificate depends on conditions and/or is limited to the following: \